


# Function description installation water detector WM1 - WM12

Function desc. WM1-WM12 902000015 GB 04/17



## Installation options:

The person or company responsible for the system must secure that:

- Safety and operation instructions are accessible and followed.
- Local accident prevention regulations and standards are obeyed.
- Performance data and installation specifications are regarded.
- Safety devices are installed and recommended maintenance is performed.
- National regulations for disposal of electrical equipment are obeyed.

## Installation and start-up

During Installation precaution are to made that the water indicator connection with mark X is installed at the bottom of the tank.

Depending on the execution, the water indicator is capable of indicating the smallest water quantum up <2% water content in oil, but due to the physical difference between water and oil, it is recommended that the container has be in a static installation position for the water indicator to be able to obtain accurate detection.



**GOLDAMMER**  
REGELUNGSTECHNIK GMBH

SCHÖLLERSHEIDER STR. 15 PHONE (49) 2104/12093  
P.O.BOX 10 02 17 FAX (49) 2104/12028  
D-40802 METTMANN [www.Goldammer-Regelungstechnik.com](http://www.Goldammer-Regelungstechnik.com)  
[info@goldammer-regelungstechnik.com](mailto:info@goldammer-regelungstechnik.com)